

## Multiparameter Photometer for Water & Wastewater





## Dear<br/>Customer,Thank you for choosing a Hanna Instruments product.<br/>Please read this instruction manual carefully before using this instrument.<br/>This manual will provide you with the necessary information for correct use of this<br/>instrument, as well as a precise idea of its versatility.<br/>If you need additional technical information, do not hesitate to e-mail us at<br/>tech@hannainst.com or view our worldwide contact list at www.hannainst.com.

All rights are reserved. Reproduction in whole or in part is prohibited without the written consent of the copyright owner, Hanna Instruments Inc., Woonsocket, Rhode Island, 02895, USA.

## 3. SPECIFICATIONS

| Measurement Channels         |                                     | 5 x optical channels<br>1 x digital electrode channel (pH measurement)   |
|------------------------------|-------------------------------------|--|
| Photometer                   | Range                               | 0.000 to 4.000 Abs   |
|                              | Resolution                          | 0.001 Abs  |
|                              | Accuracy                            | ±0.003 Abs @ 1.000 Abs   |
|                              | Light source                        | Light Emitting Diode   |
|                              | Bandpass filter bandwidth           | 8 nm   |
|                              | Bandpass filter wavelength accuracy | ±1.0 nm  |
|                              | Light detector                      | Silicon photocell  |
|                              | Cuvette types                       | Round, 24.6 mm diameter  |
|                              | Number of methods                   | 68   |
| Probe                        | Range                               | -2.00 to 16.00 pH ( $\pm$ 1000.0 mV)*  |
|                              | Resolution                          | 0.01 pH (0.1 mV)   |
|                              | Accuracy                            | ±0.01 pH (±0.2 mV) @ 25 °C / 77 °F   |
|                              | Temperature compensation            | ATC, -5.0 to 100.0 °C (23.0 to 212.0 °F)*  |
|                              | Calibration                         | two-point, from five available buffers<br>(4.01, 6.86, 7.01, 9.18, 10.01 pH)   |
|                              | Electrode                           | Intelligent pH / temperature electrode   |
| Temperature                  | Range                               | -20.0 to 120.0 °C (-4.0 to 248.0 °F)   |
|                              | Resolution                          | 0.1 °C (0.1 °F)  |
|                              | Accuracy                            | ±0.5 °C @ 25 °C (±0.9 °F @ 77 °F)  |
| Additional<br>Specifications | Logging                             | 1000 readings (mixed photometer and electrode)   |
|                              | Display                             | 128 x 64 pixel B/W LCD with backlight  |
|                              | USB-A (Host) functions              | Mass-storage host  |
|                              | USB-B (Device) functions            | Power input, mass-storage device   |
|                              | Battery life                        | > 500 photometer measurements or 50 hours of continuous pH measurement   |
|                              | Power supply                        | 5 Vdc USB 2.0 power adapter / type micro-B<br>connector<br>3.7 Vdc Li-polymer rechargeable battery,<br>non-serviceable |
|                              | Environment                         | 0 to 50 °C (32 to 122 °F)<br>0 to 95% RH, non-serviceable  |
|                              | Dimensions                          | 206 x 177 x 97 mm (8.1 x 7.0 x 3.8")   |
|                              | Weight                              | 1.0 kg (2.2 lbs.)  |

\*Limits will be reduced to actual probe / sensor limits.